

# Novembermøte 2018

20.11.18 Helse Midt RHF

# DoseTrack

- \* Oppstart 10. oktober 2017
- \* En del arbeid å legge inn modalitetene – jobba tett med medisinsk teknisk og leverandør
- \* Har kobla opp 80 modaliteter
- \* Oppdager stadig nye bruksområder

# Analyseverktøy

**SECTRA**

Search by Forenames, Surname, Date of Birth or Identifier



## Analytics

Bookmarks  + ×

Search Back Clear

Filters 157 / 1117

× Date = 10 Oct 2017 - 10 Nov 2017

× Modality = CT

Score Card Graphs Study List Exposure List Analytics Export

AET	Type	# Modalities	# Studies	# Alerts	# Rejects	% Rejects	CtDIVol Max Q3 (mGy)	DLP Tot Q3 (n
CT54514	CT	1	15	0	0	0	19.87	1048
CTAWP64151	CT	1	48	0	0	0	60.77	943.1
CTAWP73111	CT	1	16	0	0	0	16.82	1091
NMPETCTGA	CT	1	4	0	0	0	13.71	939.6
TRCT02G	CT	1	73	0	0	0	23.98	637.9
TRCTGA004	CT	1	1	0	0	0	0.1620	129.2

# Kan eksportere data

Study Level ▼

Patient	Study	Dose	Mammography Dose
<input type="checkbox"/> Patient ID	<input type="checkbox"/> Study Date	<input type="checkbox"/> Exposure Count	<input type="checkbox"/> AGD Total (mGy)
<input type="checkbox"/> Age (Years)	<input type="checkbox"/> Accession Number	<input type="checkbox"/> Protocol Name	<input type="checkbox"/> AGD Total Left (mGy)
<input type="checkbox"/> Sex	<input type="checkbox"/> Exam Code	<input type="checkbox"/> Protocol Code	<input type="checkbox"/> AGD Total Right (mGy)
<input type="checkbox"/> Height (cm)	<input type="checkbox"/> Exam Description	<input type="checkbox"/> Protocol Description	<input type="checkbox"/> Breast Tissue Density Combined (%)
<input type="checkbox"/> Weight (kg)	<input type="checkbox"/> Location	<b>Conventional Dose</b>	<input type="checkbox"/> Breast Tissue Density Left (%)
	<input type="checkbox"/> Hospital	<input type="checkbox"/> Entrance Dose (mGy)	<input type="checkbox"/> Breast Tissue Density Right (%)
	<input type="checkbox"/> Referrer Code	<input type="checkbox"/> DAP Total (Gy*cm2)	<input type="checkbox"/> Tissue Volume Left (cm3)
	<input type="checkbox"/> Referrer Description	<input type="checkbox"/> Plane Code	<input type="checkbox"/> Tissue Volume Right (cm3)
	<input type="checkbox"/> Operator	<b>CT Dose</b>	<b>Fluoroscopy Dose</b>
	<input type="checkbox"/> Performing Physician	<input type="checkbox"/> DLP Total (mGy*cm)	<input type="checkbox"/> CAK (mGy)
	<input type="checkbox"/> Pregnancy Status Code	<input type="checkbox"/> CTDIvol Max (mGy)	<input type="checkbox"/> Total Time Fluoroscopy (s)
	<input type="checkbox"/> Pregnancy Status Description	<input type="checkbox"/> CTDIvol Spiral Max (mGy)	<b>Nuc Med Dose</b>
	<input type="checkbox"/> Modality Room	<input type="checkbox"/> SSDE Effective Diameter Source	<input type="checkbox"/> Activity Total
	<input type="checkbox"/> Modality Type	<input type="checkbox"/> SSDE Effective Diameter (cm)	
	<input type="checkbox"/> Image Quality Code	<input type="checkbox"/> SSDE Coefficient	
	<input type="checkbox"/> Image Quality Description	<input type="checkbox"/> SSDE Max (mGy)	
	<input type="checkbox"/> Image Quality Reason Code		
	<input type="checkbox"/> Image Quality Reason Description		
	<input type="checkbox"/> Equipment Name		
	<input type="checkbox"/> Station Name		
	<input type="checkbox"/> AET		
	<input type="checkbox"/> Dose Alarm		
	<input type="checkbox"/> Investigation Status		
	<input type="checkbox"/> Investigation Comment		

- Patient**
- Patient ID
  - Age (Years)
  - Sex
  - Height (cm)
  - Weight (kg)
- Study**
- Study Date
  - Accession Number
  - Exam Code
  - Exam Description
  - Location
  - Hospital
  - Referrer Code
  - Referrer Description
  - Operator
  - Performing Physician
  - Pregnancy Status Code
  - Pregnancy Status Description
  - Modality Room
  - Modality Type
  - Image Quality Code
  - Image Quality Description
  - Image Quality Reason Code
  - Image Quality Reason Description
  - Equipment Name
  - Station Name
  - AET
  - Dose Alarm
  - Investigation Status
  - Investigation Comment

- Dose**
- Exposure Count
  - Protocol Name
  - Protocol Code
  - Protocol Description
- Conventional Dose**
- Entrance Dose (mGy)
  - DAP Total (Gy\*cm2)
  - Plane Code
- CT Dose**
- DLP Total (mGy\*cm)
  - CTDIVol Max (mGy)
  - CTDIVol Spiral Max (mGy)
  - SSDE Effective Diameter Source
  - SSDE Effective Diameter (cm)
  - SSDE Coefficient
  - SSDE Max (mGy)

Exposure Level ▾

- Mammography Dose**
- AGD Total (mGy)
  - AGD Total Left (mGy)
  - AGD Total Right (mGy)
  - Breast Tissue Density Combined (%)
  - Breast Tissue Density Left (%)
  - Breast Tissue Density Right (%)
  - Tissue Volume Left (cm3)
  - Tissue Volume Right (cm3)
- Fluoroscopy Dose**
- CAK (mGy)
  - Total Time Fluoroscopy (s)
- Nuc Med Dose**
- Activity Total


- Exposure**
- Ordinal
  - Acquisition Protocol
  - Acquisition Type
  - Exposure Time (ms)
  - mAs (mAs)
  - Filter Material
  - Filter Type
  - Filter Thickness
  - Distance Source to Detector (mm)
  - Tube Current (uA)
  - Tube Voltage Peak (kV)
  - Reject Reason Code
  - Target Region
- Conventional Exposure**
- AEC Used
  - DAP (Gy\*mc2)
  - Grid Code
  - Grid Description
  - Projection Code
  - Projection Description
  - Exposure Index
  - Laterality

- CT Exposure**
- CTDIVol (mGy)
  - DLP (mGy\*cm)
  - Effective Dose 103 (mSv)
  - Exposure Time Per Rotation (ms)
  - Nominal Total Collimation Width (mm)
  - Phantom Code
  - Phantom Description
  - Pitch
  - Scanning Length (mm)
  - X-Ray Source Identifier
  - SSDE (mGy)
- Mammography Exposure**
- AGD (mGy)
  - Compression Force (N)
  - Applied Pressure (kPa)
  - Compression Thickness (mm)

- Fluoroscopy Exposure**
- Collimated Field Area (m2)
  - Distance Source To Isocenter (mm)
  - Patient Orientation
  - Patient Table Relationship
  - Positioner Primary Angle (deg)
  - Positioner Secondary Angle (deg)
  - Air Kerma (mGy)
  - Pulse Rate (pulse/s)
  - Pulse Width (ms)
  - Table Cradle Tilt Angle (deg)
  - Table Head Tilt Angle (deg)
  - Table Height Position (mm)
  - Table Horizontal Rotation Angle (deg)
  - Table Lateral Position (mm)
  - Table Longitudinal Position (mm)
- Nuc Med Exposure**
- Volume
  - Activity
  - Radionuclide
  - Pharmaceutical

Export    Select All    Clear

# Scatter chart



# Går inn på enkeltundersøkelse

## Exposures

DLP **6.7** mGy-cm CTDI Volume **0.0853537** mGy

Phantom **IEC Body Dosimetry Phantom**

Time **7909** ms

Source Parameters

Tube Current **94.8286762**  $\mu$ A mAs **1** mAs

Voltage Peak **120** kV

Acq. Protocol **THX**

DLP **289.415491** mGy-cm CTDI Volume **8.5599293506** mGy

Phantom **IEC Body Dosimetry Phantom**

Time **7118** ms

Source Parameters

Tube Current **18329.7809**  $\mu$ A mAs **130** mAs

Voltage Peak **120** kV

Acq. Protocol **ABDOMEN**

DLP **880.433498** mGy-cm CTDI Volume **14.2066697884** mGy

Phantom **IEC Body Dosimetry Phantom**


Time **13047** ms

Source Parameters

Tube Current **16645.3374**  $\mu$ A mAs **217** mAs


Voltage Peak **120** kV

# Box plot






# Histogram chart




# Aldersfordeling rtg hofte







# Alarmer

## DoseTrack Alarm for Sykehuset Levanger - LECT

 no-reply@dosetrack-no.sectra.com

 Follow up. Fullført 5. november 2018.  
Ekstra linjeskift ble fjernet i denne meldingen.

Sendt: fr 02.11.2018 12:30

Til:  Gårseth, Mari

---


A dose alarm has been raised for location Sykehuset Levanger - Avdeling for bildediagnostikk, Sykehuset Levanger - LECT

Modality: LECT001 - LECT001


Procedure: CT Caput - CT Caput | CT Caput - CT Caput Accession number: NORLE00011364292

Reason: The measurement for DLP Total was 2301 mGy\*cm which is greater than 1900 mGy\*cm set as the limit Please login to the DoseTrack to assess the alarm.

# VirtualDose



## Organs vs. Dose



Organ Dose	
Organ/Tissue Name	Doses ( mGy )
Bone Surface	3.34
Brain	0.86
Breast	4.23
Colon	5.50
Esophagus	4.94
Liver	5.31
Lung	5.39
Ovaries	4.76
Red Bone Marrow	3.33
Salivary Glands	5.79
Skin	2.73
Stomach	5.30
Thyroid	6.94
Urinary Bladder	5.26
<b>Total Effective Dose(ICRP103 )</b>	
<b>(mSv): 4.86</b>	

Remainder Organs	
Remainder Organs	Doses ( mGy )
Adrenals	4.66
Eye Lens	0.55
Eyeballs	0.64
Gall Bladder	5.10
Heart	5.51
Kidneys	5.71
Muscle	3.51
Oral Mucosa	4.64
Pancreas	4.91

# Hva har vi brukt DoseTrack til?

- \* Optimalisering
- \* Etablering av representative doser
- \* Oppfølging av alarmer
  - \* Samarbeid med fagradiografer og radiologer
  - \* Internundervisning
- \* Beregning av fosterdose
- \* Identifisering av «uteliggere»

# Et par utfordringer

- \* Undersøkelser fra en nytilkoblet modalitet havnet på en annen modalitet i DoseTrack!
  - \* Årsak: Modalitetene hadde samme Station name
  - \* Løsning: Endret Station name på ene modaliteten
- \* DAP-meter ved modalitet defekt og sendte inn veldig små DAP-verdier
  - \* Ble oppdaget ved statuskontroll (ble ikke oppdaget ved «Stalled modalities» da den faktisk sendte noe...)
  - \* Løsning: Skifte DAP meter til kr 100.000,-

# Fortsettelse

- \* Webinar med SECTRA
- \* Samarbeidsmøter regionalt
- \* ... og etterhvert også nasjonalt